

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product

POWER SUPPLY

Name and address of the applicant

SL POWER ELECTRONICS CORP  
BLDG A 6050 KING DR VENTURA CA 93003 USA

Name and address of the manufacturer

SL POWER ELECTRONICS CORP  
BLDG A 6050 KING DR VENTURA CA 93003 USA

Name and address of the factory

*Note: When more than one factory, please report on page 2*SL XIANGHE POWER ELECTRONICS CORP  
No.B-02-03, North side of Landscape Ave, Qibu Distr  
Environmental Industrial Park Xianghe, Hebei 065400  
China Additional Information on page 2

Ratings and principal characteristics

Input: 100-240Vac, 50-60Hz, 1.5A

Trademark / Brand (if any)

N/A

Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

SLB65SXXYZZ  
See Page 2

Additional information (if necessary may also be reported on page 2)

Class I  
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E511446-D1021-1/A0/C0-CB issued on 2020-12-28

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2021-01-07

Signature:

Jolanta M. Wroblewska



**Model Details:**

SLB65SXXYZZ Where XX represents the output voltage which may be any number from 05 to 48. Y can be K (for Class I); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

**Factories:**

INDUSTRIAS S L S A DE C V  
CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI 21254 MEXICALI BC  
MEXICO

**Additional Information:**

**Output:**

5V/8.0A to 48V/1.36A, Max. 65W. Standard models output rating are listed as below:

For SLB65S05XXYZZ: 5V/8.0A, Max. 40W

For SLB65S12XXYZZ: 12V/5.0A, Max. 60W

For SLB65S24XXYZZ: 24V/2.71A, Max. 65W

For SLB65S48XXYZZ: 48V/1.36A, Max. 65W

Additionally evaluated to EN 60601-1:2006/A1:2013/A12:2014. National differences specified in the CB Test Report. Risk Management was not included in this investigation.

**Additional information (if necessary)**



- 
- 
- 
- 

UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2021-01-07

Signature:

Jolanta M. Wroblewska