



Ref. Certif. No.

**US-37161-UL**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

### CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

SL POWER ELECTRONICS CORP  
BLDG A  
6050 KING DR  
VENTURA, CA 93003 USA

Name and address of the manufacturer

SL POWER ELECTRONICS CORP  
BLDG A  
6050 KING DR  
VENTURA, CA 93003 USA

Name and address of the factory

Note: When more than one factory, please report on page 2

SL XIANGHE POWER ELECTRONICS CORP  
No. B-02-03, North side of Landscape Ave, Qibu  
Distr Environmental Industrial Park  
Xianghe, Hebei 065400 China

Additional Information on page 2

Ratings and principal characteristics

Input: 100-240Vac, 50-60Hz, 8.0A,

Output: See CB Test Report for details

Trademark / Brand (if any)

N/A

Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

NGB660SXXYZZ  
See Page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 60601-1:2006 / A1:2013 / A12:2014; National Differences specified in the CB Test Report.

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E511446-D1018-1/A0/C0-CB issued on 2020-11-24

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2020-12-08

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

**US-37161-UL**

Model Details:  
NGB660SXXYZZ Where XX represents the output voltage which may be any number from 12 to 48. Y is K (for Class I); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

Factories:  
INDUSTRIAS S L S A DE C V  
CIRCUITO SIGLO XXI 2055 COL PARQUE  
INDUSTRIAL EX-XXI  
21254 MEXICALI BC Mexico

Additional Information:  
The risk management requirements of the standard were not addressed.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2020-12-08

Signature:

Jolanta M. Wroblewska