



Ref. Certif. No.

US-37162-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

SL POWER ELECTRONICS CORP
BLDG A
6050 KING DR
VENTURA, CA 93003 USA

Name and address of the manufacturer

SL POWER ELECTRONICS CORP
BLDG A
6050 KING DR
VENTURA, CA 93003 USA

Name and address of the factory

Note: When more than one factory, please report on page 2

SL XIANGHE POWER ELECTRONICS CORP
No. B-02-03, North side of Landscape Ave, Qibu Distr
Environmental Industrial Park
Xianghe, Hebei 065400 China

Additional Information on page 2

Ratings and principal characteristics

Input: 100-240Vac, 50-60Hz, 8.0A,
Output: See CB Test Report for details

Trademark / Brand (if any)

N/A

Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

NGB660SXXYZZ
See Page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 60601-1:2006 / A1:2013 / A12:2014; National Differences specified in the CB Test Report.

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E511446-D1018-1/A0/C0-CB issued on 2020-11-24

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2020-12-08

Signature:

Jolanta M. Wroblewska

Model Details:

NGB660SXXYZZ Where XX represents the output voltage which may be any number from 12 to 48. Y is C (for Class II); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

Factories:

INDUSTRIAS S L S A DE C V
CIRCUITO SIGLO XXI 2055 COL PARQUE
INDUSTRIAL EX-XXI
21254 MEXICALI BC Mexico

Additional Information:

The risk management requirements of the standard were not addressed.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2020-12-08

Signature:

Jolanta M. Wroblewska