

Certificate



Certificate no.

CU 72150542 04

License Holder:

Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
Little Island, Cork
Ireland

Manufacturing Plant:

Shenzhen WATT Electronics Co., Ltd.
No. 5, Tunnel 1, Tangfang Garden
35th District, Baoan County
518101 Shenzhen, Guangdong
China

Test report no.: USA-JAK 31482538 005

Client Reference: Diarmuid Hogan

Tested to:

UL 60950-1:2007 R10.14

CAN/CSA-C22.2 NO.60950-1-07+A1:2011+A2:2014

Certified Product: Switched Mode Power Supply

License Fee - Units


Change:

Output Ratings DC:

- 1) Output 1: 14-28V/42A (24V/42A nominal)
- 2) Output 1: 14-28V/21A (24V/21A nominal)
- 3) Output 1: 29-58V/21A (48V/21A nominal)
- 4) Output 1: 29-58V/10.5A (48V/10.5A nominal)
- 5) Output 1: 19-40V/28A (36V/28A nominal)
- 6) Output 1: 19-40V/14A (36V/14A nominal)

Appendix: 1, 1-9

Licensed Test mark:



Date of Issue

(day/mo/yr)

27/06/2016