

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switching power supply unit
(Switch Mode Power Supply for Building-in)

Name and address of the applicant
Nom et adresse du demandeur

Astec International Ltd.
16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong
Kowloon, HONG KONG

Name and address of the manufacturer
Nom et adresse du fabricant

Astec International Ltd., 16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun
Tong, Kowloon, HONG KONG

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc., 104 Laguna Boulevard, Laguna Technopark,
4026 Sta. Rosa Laguna, PHILIPPINES
For further information please see attachment

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Rated Input : 100-250VAC, 2.5A, 50/60Hz or
140-300VDC, 2.5A
Rated Output : +5VDC, 11.0A MAX (for model NPS62-M)
+12VDC, 5.0A MAX (for model NPS63-M)
+24VDC, 2.5A MAX (for model NPS65-M)
+15VDC, 4.0A MAX (for model NPS64-M)
Maximum output power:
For models NPS63-M, NPS65-M
and NPS64-M:
60W with convection cooling
60W with forced air cooling
For model NPS62-M:
55W with convection cooling
55W with forced air cooling

Model/type Ref.
Ref. de type

NPS62-M, NPS63-M, NPS65-M, NPS64-M

Additional information (if necessary)
Information complémentaire (si nécessaire)

See Test Report for National Differences and Group Differences

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3);am1

as shown in the Test Report Ref. No.
which forms part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificat

TÜV SÜD PSB Pte Ltd
211-400954-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification

Date, 2015-05-19
CBS 15 05 13890 02221


(Yager Bi)



TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

PSB Singapore

Additional factory information:

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc.,
Main Avenue Corner Rd "J",
Cavite Export Processing Zone, Rosario,
Cavite, Philippines

Zhongshan General Carton Box Factory Co., Ltd.
No.62, Qi Guan Road West, Shiqi District, 528400,
Zhongshan City, Guangdong,
People's Republic of China

Astec Electronics (Luoding) Co. Ltd.
No. 68 Baocheng Road East, Fucheng Luoding,
527200 Guangdong,
People's Republic of China

Additional Information:

*The risk management requirements of the standard
were not addressed.*

Date, 2015-05-19
CBS 15 05 13890 02221

Signature : (Yager Bi)

