



Ref. Certif. No.

US-45106-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Component - Power Supply
Name and address of the applicant	SL POWER ELECTRONICS CORP 1595 WYNKOOP STREET SUITE 800 DENVER, CO 80202 UNITED STATES
Name and address of the manufacturer	SL POWER ELECTRONICS CORP 1595 WYNKOOP STREET SUITE 800 DENVER, CO 80202 UNITED STATES
Name and address of the factory	INDUSTRIAS S L S A DE C V CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI MEXICALI, BAJA CALIFORNIA 21254 MEXICO <input type="checkbox"/> Additional Information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	Input: 100-240Vac, 50-60Hz, 1.3A Output: See Test Report Enclosure - Miscellaneous for examples of models / ratings
Trademark / Brand (if any)	 
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	MINT1065AX75CZ, MINT1065DX75YZ <input checked="" type="checkbox"/> Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	The risk management requirements of the standard were not addressed National Differences: EU Group Differences, CA, JP, US <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E116994-D6015-CB-1 issued on 2025-04-16

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2025-04-16

Signature: Mauricio Avila



Ref. Certif. No.

US-45106-UL

Additional Model Detail(s):

MINT1065AX75CZ,

where X = any number 12 through 48, and Z is any number 01 through 99, designates additional configurations indicating non-safety related options.

MINT1065DX75YZ,

where X = any number 12 through 48, Y = C or K, and Z = any number 01 through 99, designates additional configurations indicating non-safety related options

Additionally evaluated to:

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021

Additional information (if necessary)



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2025-04-16

Signature:

Mauricio Avila